

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042221\
 Data File : VX021831.D
 Acq On : 22 Apr 2021 20:44
 Operator : JC/MD
 Sample : M2138-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-MW-1S(ONSITE)

Quant Time: Apr 23 07:46:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 20 13:46:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	157704	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	289803	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.091	117	314132	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	165366	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	91824	51.33	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.66%	
35) Dibromofluoromethane	5.464	113	92816	50.08	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.16%	
50) Toluene-d8	8.695	98	360070	50.53	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.06%	
62) 4-Bromofluorobenzene	11.116	95	158607	59.60	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 119.20%	
Target Compounds						
16) Acetone	2.422	43	1596	2.59	ug/l	91
64) Tetrachloroethene	9.317	164	1720	0.70	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042221\
Data File : VX021831.D
Acq On : 22 Apr 2021 20:44
Operator : JC/MD
Sample : M2138-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-MW-1S(ONSITE)

Quant Time: Apr 23 07:46:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 20 13:46:22 2021
Response via : Initial Calibration

