

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042222\
 Data File : VX028278.D
 Acq On : 22 Apr 2022 14:52
 Operator : JC/MD
 Sample : N2511-07 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1502

Quant Time: Apr 23 06:41:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	450334	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	753612	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	721569	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	359839	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	299018	50.924	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 101.840%	
35) Dibromofluoromethane	5.385	113	278763	52.746	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 105.500%	
50) Toluene-d8	8.653	98	1019713	50.235	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 100.480%	
62) 4-Bromofluorobenzene	11.085	95	388671	50.078	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 100.160%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	22637	11.289	ug/l	95
18) Methyl Acetate	2.709	43	2870	0.550	ug/l #	88
25) 2-Butanone	4.574	43	3911	1.125	ug/l	92
40) Benzene	6.037	78	4715	0.234	ug/l	99
52) Toluene	8.720	92	3200	0.228	ug/l	99
95) Naphthalene	13.780	128	8759	0.368	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042222\
Data File : VX028278.D
Acq On : 22 Apr 2022 14:52
Operator : JC/MD
Sample : N2511-07 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1502

Quant Time: Apr 23 06:41:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

