

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042222\
 Data File : VX028283.D
 Acq On : 22 Apr 2022 16:49
 Operator : JC/MD
 Sample : N2511-13 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1553

Quant Time: Apr 23 06:43:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	451269	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	747290	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	700522	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	337087	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	302574	51.423	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.840%	
35) Dibromofluoromethane	5.391	113	284508	54.288	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 108.580%	
50) Toluene-d8	8.653	98	1017260	50.538	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 101.080%	
62) 4-Bromofluorobenzene	11.079	95	369395	47.997	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 96.000%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	8243	3.209	ug/l	99
20) Methylene Chloride	2.800	84	1089	0.249	ug/l #	85
25) 2-Butanone	4.562	43	2725	0.782	ug/l #	80
52) Toluene	8.720	92	4611	0.332	ug/l	98
68) m/p-Xylenes	10.305	106	5171	0.507	ug/l	92
69) o-Xylene	10.646	106	3702	0.364	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	6375	0.324	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	29140	1.491	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042222\
Data File : VX028283.D
Acq On : 22 Apr 2022 16:49
Operator : JC/MD
Sample : N2511-13 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1553

Quant Time: Apr 23 06:43:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

