

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042222\
 Data File : VX028284.D
 Acq On : 22 Apr 2022 17:12
 Operator : JC/MD
 Sample : N2519-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Apr 23 06:43:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	439305	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	737525	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	696795	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	333654	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	302820	52.866	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	105.740%
35) Dibromofluoromethane	5.391	113	279063	53.954	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	107.900%
50) Toluene-d8	8.653	98	1024832	51.589	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.180%
62) 4-Bromofluorobenzene	11.079	95	374375	49.288	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.580%
Target Compounds						
20) Methylene Chloride	2.788	84	1492	0.351	ug/l	Qvalue # 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042222\
Data File : VX028284.D
Acq On : 22 Apr 2022 17:12
Operator : JC/MD
Sample : N2519-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Apr 23 06:43:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

