

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042222\
 Data File : VX028285.D
 Acq On : 22 Apr 2022 17:35
 Operator : JC/MD
 Sample : N2521-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB

Quant Time: Apr 23 06:43:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	435124	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	729628	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	684978	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	314741	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	299161	52.729	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	105.460%
35) Dibromofluoromethane	5.385	113	274557	53.658	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	107.320%
50) Toluene-d8	8.653	98	1007364	51.258	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.520%
62) 4-Bromofluorobenzene	11.079	95	359053	47.782	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042222\
Data File : VX028285.D
Acq On : 22 Apr 2022 17:35
Operator : JC/MD
Sample : N2521-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB

Quant Time: Apr 23 06:43:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

