

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042222\
 Data File : VX028286.D
 Acq On : 22 Apr 2022 17:59
 Operator : JC/MD
 Sample : N2511-06 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1530

Quant Time: Apr 23 06:44:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	429736	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	731798	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	694559	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	340503	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	293417	52.365	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 104.740%	
35) Dibromofluoromethane	5.391	113	270402	52.689	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 105.380%	
50) Toluene-d8	8.653	98	989240	50.187	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 100.380%	
62) 4-Bromofluorobenzene	11.079	95	365904	48.549	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 97.100%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	26125	13.947	ug/l	97
25) 2-Butanone	4.556	43	51748	15.594	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042222\
Data File : VX028286.D
Acq On : 22 Apr 2022 17:59
Operator : JC/MD
Sample : N2511-06 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1530

Quant Time: Apr 23 06:44:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

