

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042222\
 Data File : VX028287.D
 Acq On : 22 Apr 2022 18:22
 Operator : JC/MD
 Sample : N2519-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-1

Quant Time: Apr 23 06:44:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	453548	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	761318	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	719041	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	341823	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	300398	50.797	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.600%
35) Dibromofluoromethane	5.391	113	280745	52.583	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	105.160%
50) Toluene-d8	8.653	98	1015961	49.544	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.080%
62) 4-Bromofluorobenzene	11.085	95	374815	47.803	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.600%
Target Compounds						
						Qvalue
20) Methylene Chloride	2.794	84	1520	0.346	ug/l	# 77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042222\
Data File : VX028287.D
Acq On : 22 Apr 2022 18:22
Operator : JC/MD
Sample : N2519-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-1

Quant Time: Apr 23 06:44:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

