

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042222\
 Data File : VX028293.D
 Acq On : 22 Apr 2022 20:41
 Operator : JC/MD
 Sample : N2511-10 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1593

Quant Time: Apr 23 06:45:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	422607	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	727123	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	700073	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	334829	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	296678	53.841	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	107.680%
35) Dibromofluoromethane	5.391	113	272273	53.395	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	106.780%
50) Toluene-d8	8.653	98	994923	50.800	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.600%
62) 4-Bromofluorobenzene	11.085	95	368366	49.190	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.380%
Target Compounds						
						Qvalue
16) Acetone	2.379	43	6633	2.559	ug/l	88
20) Methylene Chloride	2.800	84	1403	0.343	ug/l #	63
25) 2-Butanone	4.568	43	11636	3.566	ug/l	100
52) Toluene	8.720	92	329983	24.395	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042222\
Data File : VX028293.D
Acq On : 22 Apr 2022 20:41
Operator : JC/MD
Sample : N2511-10 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1593

Quant Time: Apr 23 06:45:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

