

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042222\
 Data File : VX028300.D
 Acq On : 23 Apr 2022 00:10
 Operator : JC/MD
 Sample : PB144273ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144273ZHE#01

Quant Time: Apr 23 06:48:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	429824	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	734591	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	728808	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	341961	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	252387	45.034	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	90.060%
35) Dibromofluoromethane	5.391	113	233957	45.414	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	90.820%
50) Toluene-d8	8.653	98	876661	44.306	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	88.620%
62) 4-Bromofluorobenzene	11.079	95	325420	43.014	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	86.020%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	11049	5.087	ug/l	98
20) Methylene Chloride	2.788	84	24214	5.823	ug/l	98
43) Isopropyl Acetate	6.342	43	29754	2.681	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042222\
Data File : VX028300.D
Acq On : 23 Apr 2022 00:10
Operator : JC/MD
Sample : PB144273ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144273ZHE#01

Quant Time: Apr 23 06:48:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

