

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042222\
 Data File : VX028301.D
 Acq On : 23 Apr 2022 00:33
 Operator : JC/MD
 Sample : PB144273ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144273ZHE#02

Quant Time: Apr 23 06:48:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	419559	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	714492	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	708576	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	332682	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	247139	45.176	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	90.360%
35) Dibromofluoromethane	5.391	113	229497	45.802	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	91.600%
50) Toluene-d8	8.653	98	850904	44.214	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	88.420%
62) 4-Bromofluorobenzene	11.079	95	316687	43.037	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	86.080%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	11698	5.637	ug/l #	86
20) Methylene Chloride	2.788	84	26818	6.607	ug/l	98
43) Isopropyl Acetate	6.348	43	26200	2.427	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042222\
Data File : VX028301.D
Acq On : 23 Apr 2022 00:33
Operator : JC/MD
Sample : PB144273ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144273ZHE#02

Quant Time: Apr 23 06:48:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

