

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042222\
 Data File : VX028302.D
 Acq On : 23 Apr 2022 00:56
 Operator : JC/MD
 Sample : PB144273ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144273ZHE#03

Quant Time: Apr 23 06:49:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	410254	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	709746	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	719909	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	336944	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	246420	46.066	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.140%
35) Dibromofluoromethane	5.391	113	227469	45.700	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	91.400%
50) Toluene-d8	8.652	98	845728	44.239	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	88.480%
62) 4-Bromofluorobenzene	11.079	95	321618	43.999	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.000%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	12119	6.055	ug/l	98
20) Methylene Chloride	2.794	84	26980	6.798	ug/l	99
43) Isopropyl Acetate	6.348	43	26729	2.493	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042222\
Data File : VX028302.D
Acq On : 23 Apr 2022 00:56
Operator : JC/MD
Sample : PB144273ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144273ZHE#03

Quant Time: Apr 23 06:49:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

