

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042222\
 Data File : VX028305.D
 Acq On : 23 Apr 2022 02:06
 Operator : JC/MD
 Sample : PB144273ZHE#06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144273ZHE#06

Quant Time: Apr 23 06:50:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	411143	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	703354	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	706590	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	326416	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	246140	45.915	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.820%
35) Dibromofluoromethane	5.391	113	225808	45.779	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	91.560%
50) Toluene-d8	8.653	98	849132	44.821	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	89.640%
62) 4-Bromofluorobenzene	11.079	95	315039	43.491	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	86.980%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	11991	5.961	ug/l	96
20) Methylene Chloride	2.788	84	27354	6.877	ug/l	97
43) Isopropyl Acetate	6.349	43	27170	2.557	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042222\
Data File : VX028305.D
Acq On : 23 Apr 2022 02:06
Operator : JC/MD
Sample : PB144273ZHE#06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144273ZHE#06

Quant Time: Apr 23 06:50:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

