

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042222\
 Data File : VX028308.D
 Acq On : 23 Apr 2022 03:16
 Operator : JC/MD
 Sample : PB144273ZHE#09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144273ZHE#09

Quant Time: Apr 23 06:51:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	411293	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	711994	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	708496	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	317382	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	246646	45.992	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.980%
35) Dibromofluoromethane	5.391	113	226641	45.390	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	90.780%
50) Toluene-d8	8.653	98	849713	44.307	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	88.620%
62) 4-Bromofluorobenzene	11.085	95	309700	42.235	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	84.480%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	12188	6.079	ug/l	96
20) Methylene Chloride	2.788	84	25751	6.472	ug/l	99
43) Isopropyl Acetate	6.348	43	28155	2.618	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042222\
Data File : VX028308.D
Acq On : 23 Apr 2022 03:16
Operator : JC/MD
Sample : PB144273ZHE#09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144273ZHE#09

Quant Time: Apr 23 06:51:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

