

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042222\
 Data File : VX028310.D
 Acq On : 23 Apr 2022 04:02
 Operator : JC/MD
 Sample : N2521-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW1

Quant Time: Apr 23 06:51:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	405607	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	690129	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	654333	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	301015	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	291111	55.045	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	110.080%
35) Dibromofluoromethane	5.391	113	263240	54.390	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	108.780%
50) Toluene-d8	8.653	98	969861	52.175	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.340%
62) 4-Bromofluorobenzene	11.079	95	348554	49.040	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.080%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	6608	1.616	ug/l	95
18) Methyl Acetate	2.709	43	1450	0.308	ug/l #	80
19) Methyl tert-butyl Ether	3.117	73	30255	2.328	ug/l	95
27) cis-1,2-Dichloroethene	4.501	96	7585	1.501	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042222\
Data File : VX028310.D
Acq On : 23 Apr 2022 04:02
Operator : JC/MD
Sample : N2521-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW1

Quant Time: Apr 23 06:51:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

