

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042222\
 Data File : VX028312.D
 Acq On : 23 Apr 2022 04:49
 Operator : JC/MD
 Sample : N2521-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW2

Quant Time: Apr 23 06:52:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	403994	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	703074	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	676513	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	319012	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	289038	54.871	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 109.740%	
35) Dibromofluoromethane	5.391	113	262458	53.230	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 106.460%	
50) Toluene-d8	8.653	98	857515	45.281	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 90.560%	
62) 4-Bromofluorobenzene	11.079	95	348748	48.164	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 96.320%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	1296	0.318	ug/l	97
18) Methyl Acetate	2.709	43	1030	0.220	ug/l	90
19) Methyl tert-butyl Ether	3.117	73	18812	1.453	ug/l	99
27) cis-1,2-Dichloroethene	4.495	96	11155	2.216	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042222\
Data File : VX028312.D
Acq On : 23 Apr 2022 04:49
Operator : JC/MD
Sample : N2521-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW2

Quant Time: Apr 23 06:52:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

