

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042222\
 Data File : VX028313.D
 Acq On : 23 Apr 2022 05:12
 Operator : JC/MD
 Sample : N2519-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-4

Quant Time: Apr 23 06:52:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	387632	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	672771	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	639641	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	292243	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	284113	56.212	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 112.420%	
35) Dibromofluoromethane	5.391	113	254226	53.883	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 107.760%	
50) Toluene-d8	8.653	98	938638	51.798	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 103.600%	
62) 4-Bromofluorobenzene	11.079	95	335993	48.492	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 96.980%	
Target Compounds						
16) Acetone	2.379	43	5627	2.262	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042222\
Data File : VX028313.D
Acq On : 23 Apr 2022 05:12
Operator : JC/MD
Sample : N2519-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-4

Quant Time: Apr 23 06:52:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

