



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042224\
Data File : VX041179.D
Acq On : 22 Apr 2024 08:35
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042224W.M
Title : SW846 8260
Last Update : Tue Apr 23 05:30:28 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	17.7	9559	PASS
75	95	30	60	48.1	25989	PASS
95	95	100	100	100.0	54051	PASS
96	95	5	9	6.7	3599	PASS
173	174	0.00	2	0.5	283	PASS
174	95	50	100	96.3	52051	PASS
175	174	5	9	7.4	3845	PASS
176	174	95	101	99.1	51560	PASS
177	176	5	9	6.4	3312	PASS