

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX042225\  
 Data File : VX045881.D  
 Acq On : 22 Apr 2025 13:44  
 Operator : JC/MD  
 Sample : PB167661ZHE#02  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 8 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 PB167661ZHE#02

Quant Time: Apr 23 01:33:59 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X040225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Apr 02 03:11:43 2025  
 Response via : Initial Calibration


| Compound                      | R.T.   | QIon     | Response | Conc   | Units    | Dev(Min) |
|-------------------------------|--------|----------|----------|--------|----------|----------|
| -----                         |        |          |          |        |          |          |
| Internal Standards            |        |          |          |        |          |          |
| 1) Pentafluorobenzene         | 5.544  | 168      | 62970    | 50.000 | ug/l     | 0.00     |
| 34) 1,4-Difluorobenzene       | 6.757  | 114      | 121687   | 50.000 | ug/l     | 0.00     |
| 63) Chlorobenzene-d5          | 10.049 | 117      | 115940   | 50.000 | ug/l     | 0.00     |
| 72) 1,4-Dichlorobenzene-d4    | 12.018 | 152      | 48065    | 50.000 | ug/l     | 0.00     |
| System Monitoring Compounds   |        |          |          |        |          |          |
| 33) 1,2-Dichloroethane-d4     | 5.952  | 65       | 61095    | 53.054 | ug/l     | 0.00     |
| Spiked Amount 50.000          | Range  | 74 - 125 | Recovery | =      | 106.100% |          |
| 35) Dibromofluoromethane      | 5.379  | 113      | 44711    | 51.787 | ug/l     | 0.00     |
| Spiked Amount 50.000          | Range  | 75 - 124 | Recovery | =      | 103.580% |          |
| 50) Toluene-d8                | 8.647  | 98       | 156776   | 52.025 | ug/l     | 0.00     |
| Spiked Amount 50.000          | Range  | 86 - 113 | Recovery | =      | 104.040% |          |
| 62) 4-Bromofluorobenzene      | 11.079 | 95       | 58944    | 53.699 | ug/l     | 0.00     |
| Spiked Amount 50.000          | Range  | 77 - 121 | Recovery | =      | 107.400% |          |
| Target Compounds              |        |          |          |        |          |          |
|                               |        |          |          |        | Qvalue   |          |
| 5) Bromomethane               | 1.599  | 94       | 87       | 0.207  | ug/l #   | 72       |
| 11) Tert butyl alcohol        | 2.971  | 59       | 373      | 2.410  | ug/l #   | 86       |
| 13) Acrolein                  | 2.246  | 56       | 125      | 0.587  | ug/l #   | 66       |
| 14) Allyl chloride            | 2.660  | 41       | 304      | 0.212  | ug/l #   | 43       |
| 16) Acetone                   | 2.380  | 43       | 25046    | 52.967 | ug/l     | 99       |
| 17) Carbon Disulfide          | 2.502  | 76       | 685      | 0.367  | ug/l #   | 76       |
| 18) Methyl Acetate            | 2.709  | 43       | 1601     | 1.471  | ug/l     | 95       |
| 20) Methylene Chloride        | 2.782  | 84       | 2718     | 3.070  | ug/l     | 90       |
| 25) 2-Butanone                | 4.593  | 43       | 3697     | 5.341  | ug/l     | 91       |
| 29) Tetrahydrofuran           | 5.032  | 42       | 104      | 0.232  | ug/l #   | 35       |
| 31) Cyclohexane               | 5.538  | 56       | 1859     | 1.316  | ug/l #   | 33       |
| 36) 1,1-Dichloropropene       | 5.544  | 75       | 6285     | 5.391  | ug/l #   | 47       |
| 38) Carbon Tetrachloride      | 5.550  | 117      | 7379     | 5.857  | ug/l #   | 15       |
| 43) Isopropyl Acetate         | 6.349  | 43       | 11006    | 4.943  | ug/l     | 99       |
| 48) Methyl methacrylate       | 7.793  | 41       | 45578    | 39.647 | ug/l #   | 49       |
| 49) 1,4-Dioxane               | 7.799  | 88       | 22       | 1.058  | ug/l #   | 1        |
| 51) 4-Methyl-2-Pentanone      | 8.543  | 43       | 40446    | 27.402 | ug/l #   | 49       |
| 53) t-1,3-Dichloropropene     | 8.958  | 75       | 270      | 2.486  | ug/l #   | 43       |
| 54) cis-1,3-Dichloropropene   | 8.543  | 75       | 15303    | 11.958 | ug/l #   | 73       |
| 57) 1,3-Dichloropropane       | 9.482  | 76       | 1607     | 1.075  | ug/l #   | 42       |
| 59) 2-Hexanone                | 9.378  | 43       | 14402    | 13.176 | ug/l #   | 28       |
| 74) N-amyl acetate            | 10.896 | 43       | 401      | 0.218  | ug/l #   | 46       |
| 76) 1,2,3-Trichloropropane    | 11.079 | 75       | 34114    | 29.097 | ug/l #   | 37       |
| 81) trans-1,4-Dichloro-2-b... | 11.079 | 75       | 34114    | 98.752 | ug/l #   | 1        |
| 93) 1,2,4-Trichlorobenzene    | 13.597 | 180      | 286      | 0.319  | ug/l #   | 65       |
| 95) Naphthalene               | 13.780 | 128      | 1580     | 0.474  | ug/l #   | 94       |
| 96) 1,2,3-Trichlorobenzene    | 13.963 | 180      | 285      | 0.304  | ug/l     | 79       |
| -----                         |        |          |          |        |          |          |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX042225\  
Data File : VX045881.D  
Acq On : 22 Apr 2025 13:44  
Operator : JC/MD  
Sample : PB167661ZHE#02  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 8 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
PB167661ZHE#02

Quant Time: Apr 23 01:33:59 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X040225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Apr 02 03:11:43 2025  
Response via : Initial Calibration









#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 5.544 min Scan# 71  
Delta R.T. -0.006 min  
Lab File: VX045881.D  
Acq: 22 Apr 2025 13:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
PB167661ZHE#02


Tgt Ion:168 Resp: 62970  
Ion Ratio Lower Upper  
168 100  
99 69.4 52.3 78.5



#5  
Bromomethane  
Concen: 0.207 ug/l  
RT: 1.599 min Scan# 84  
Delta R.T. 0.006 min  
Lab File: VX045881.D  
Acq: 22 Apr 2025 13:44

Tgt Ion: 94 Resp: 87  
Ion Ratio Lower Upper  
94 100  
96 67.0 74.8 112.2#









#11  
Tert butyl alcohol  
Concen: 2.410 ug/l  
RT: 2.971 min Scan# 309  
Delta R.T. -0.006 min  
Lab File: VX045881.D  
Acq: 22 Apr 2025 13:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
PB167661ZHE#02


Tgt Ion: 59 Resp: 373  
Ion Ratio Lower Upper  
59 100  
57 5.4 8.6 12.8#



#13  
Acrolein  
Concen: 0.587 ug/l  
RT: 2.246 min Scan# 190  
Delta R.T. 0.006 min  
Lab File: VX045881.D  
Acq: 22 Apr 2025 13:44

Tgt Ion: 56 Resp: 125  
Ion Ratio Lower Upper  
56 100  
55 42.4 56.2 84.2#









#14  
 Allyl chloride  
 Concen: 0.212 ug/l  
 RT: 2.660 min Scan# 212  
 Delta R.T. 0.000 min  
 Lab File: VX045881.D  
 Acq: 22 Apr 2025 13:44




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 PB167661ZHE#02


Tgt Ion: 41 Resp: 304  
 Ion Ratio Lower Upper  
 41 100  
 39 26.3 60.2 90.2#  
 76 0.0 24.6 36.8#



#16  
 Acetone  
 Concen: 52.967 ug/l  
 RT: 2.380 min Scan# 212  
 Delta R.T. -0.006 min  
 Lab File: VX045881.D  
 Acq: 22 Apr 2025 13:44

Tgt Ion: 43 Resp: 25046  
 Ion Ratio Lower Upper  
 43 100  
 58 27.5 22.3 33.5









#17  
Carbon Disulfide  
Concen: 0.367 ug/l  
RT: 2.502 min Scan# 21  
Delta R.T. -0.006 min  
Lab File: VX045881.D  
Acq: 22 Apr 2025 13:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
PB167661ZHE#02


Tgt Ion: 76 Resp: 685  
Ion Ratio Lower Upper  
76 100  
78 0.0 7.0 10.6#



#18  
Methyl Acetate  
Concen: 1.471 ug/l  
RT: 2.709 min Scan# 266  
Delta R.T. 0.006 min  
Lab File: VX045881.D  
Acq: 22 Apr 2025 13:44

Tgt Ion: 43 Resp: 1601  
Ion Ratio Lower Upper  
43 100  
74 24.2 17.5 26.3









#20  
Methylene Chloride  
Concen: 3.070 ug/l  
RT: 2.782 min Scan# 21  
Delta R.T. 0.000 min  
Lab File: VX045881.D  
Acq: 22 Apr 2025 13:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
PB167661ZHE#02


Tgt Ion: 84 Resp: 2718  
Ion Ratio Lower Upper  
84 100  
49 146.5 107.5 161.3  
51 43.3 33.0 49.6  
86 75.0 50.1 75.1



#25  
2-Butanone  
Concen: 5.341 ug/l  
RT: 4.593 min Scan# 575  
Delta R.T. 0.037 min  
Lab File: VX045881.D  
Acq: 22 Apr 2025 13:44

Tgt Ion: 43 Resp: 3697  
Ion Ratio Lower Upper  
43 100  
72 19.1 18.6 28.0





#29

Tetrahydrofuran

Concen: 0.232 ug/l

RT: 5.032 min Scan# 64

Delta R.T. 0.031 min

Lab File: VX045881.D

Acq: 22 Apr 2025 13:44

Instrument :

MSVOA\_X

ClientSampleId :

PB167661ZHE#02





Tgt Ion: 42 Resp: 104

Ion Ratio Lower Upper

42 100

72 0.0 33.4 50.0#

71 0.0 31.4 47.0#



#31

Cyclohexane

Concen: 1.316 ug/l

RT: 5.538 min Scan# 730

Delta R.T. 0.073 min

Lab File: VX045881.D

Acq: 22 Apr 2025 13:44




Tgt Ion: 56 Resp: 1859

Ion Ratio Lower Upper


56 100

69 151.8 23.6 35.4#

84 95.0 67.7 101.5







#33

1,2-Dichloroethane-d4

Concen: 53.054 ug/l

RT: 5.952 min Scan# 798

Delta R.T. 0.000 min

Lab File: VX045881.D

Acq: 22 Apr 2025 13:44

Instrument :

MSVOA\_X

ClientSampleId :





PB167661ZHE#02

Tgt Ion: 65 Resp: 61095

Ion Ratio Lower Upper

65 100

67 49.6 0.0 99.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. 0.000 min

Lab File: VX045881.D

Acq: 22 Apr 2025 13:44




Tgt Ion: 114 Resp: 121687


Ion Ratio Lower Upper

114 100

63 24.3 0.0 46.8

88 17.8 0.0 35.4









#35  
Dibromofluoromethane  
Concen: 51.787 ug/l  
RT: 5.379 min Scan# 704  
Delta R.T. 0.000 min  
Lab File: VX045881.D  
Acq: 22 Apr 2025 13:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
PB167661ZHE#02


Tgt Ion:113 Resp: 44711  
Ion Ratio Lower Upper  
113 100  
111 102.0 81.8 122.6  
192 17.0 13.8 20.6



#36  
1,1-Dichloropropene  
Concen: 5.391 ug/l  
RT: 5.544 min Scan# 731  
Delta R.T. -0.140 min  
Lab File: VX045881.D  
Acq: 22 Apr 2025 13:44


Tgt Ion: 75 Resp: 6285  
Ion Ratio Lower Upper  
75 100  
110 5.0 16.9 50.6#  
77 0.0 25.0 37.4#








#38  
Carbon Tetrachloride  
Concen: 5.857 ug/l  
RT: 5.550 min Scan# 71  
Delta R.T. -0.122 min  
Lab File: VX045881.D  
Acq: 22 Apr 2025 13:44

Instrument :  
MSVOA\_X  
ClientSampleId :  
PB167661ZHE#02







Tgt Ion:117 Resp: 7379  
Ion Ratio Lower Upper  
117 100  
119 4.5 76.9 115.3#  
121 0.0 24.2 36.2#



#43  
Isopropyl Acetate  
Concen: 4.943 ug/l  
RT: 6.349 min Scan# 863  
Delta R.T. 0.012 min  
Lab File: VX045881.D  
Acq: 22 Apr 2025 13:44

Tgt Ion: 43 Resp: 11006  
Ion Ratio Lower Upper  
43 100  
61 18.1 14.1 21.1  
87 10.8 9.2 13.8









#48  
Methyl methacrylate  
Concen: 39.647 ug/l  
RT: 7.793 min Scan# 1100  
Delta R.T. 0.104 min  
Lab File: VX045881.D  
Acq: 22 Apr 2025 13:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
PB167661ZHE#02


Tgt Ion: 41 Resp: 45578  
Ion Ratio Lower Upper  
41 100  
69 0.0 58.2 87.2#  
39 58.0 53.6 80.4



#49  
1,4-Dioxane  
Concen: 1.058 ug/l  
RT: 7.799 min Scan# 1101  
Delta R.T. 0.140 min  
Lab File: VX045881.D  
Acq: 22 Apr 2025 13:44

Tgt Ion: 88 Resp: 22  
Ion Ratio Lower Upper  
88 100  
43 2000640.9 32.2 48.4#  
58 8722.7 58.6 88.0#









#50  
Toluene-d8  
Concen: 52.025 ug/l  
RT: 8.647 min Scan# 11  
Delta R.T. 0.000 min  
Lab File: VX045881.D  
Acq: 22 Apr 2025 13:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
PB167661ZHE#02


Tgt Ion: 98 Resp: 156776  
Ion Ratio Lower Upper  
98 100  
100 65.8 52.2 78.4



#51  
4-Methyl-2-Pentanone  
Concen: 27.402 ug/l  
RT: 8.543 min Scan# 1223  
Delta R.T. -0.030 min  
Lab File: VX045881.D  
Acq: 22 Apr 2025 13:44

Tgt Ion: 43 Resp: 40446  
Ion Ratio Lower Upper  
43 100  
58 5.6 28.5 42.7#









#53  
 t-1,3-Dichloropropene  
 Concen: 2.486 ug/l  
 RT: 8.958 min Scan# 1123  
 Delta R.T. -0.018 min  
 Lab File: VX045881.D  
 Acq: 22 Apr 2025 13:44




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 PB167661ZHE#02


Tgt Ion: 75 Resp: 270  
 Ion Ratio Lower Upper  
 75 100  
 77 0.0 25.4 38.0#



#54  
 cis-1,3-Dichloropropene  
 Concen: 11.958 ug/l  
 RT: 8.543 min Scan# 1223  
 Delta R.T. 0.177 min  
 Lab File: VX045881.D  
 Acq: 22 Apr 2025 13:44

Tgt Ion: 75 Resp: 15303  
 Ion Ratio Lower Upper  
 75 100  
 77 0.6 24.6 36.8#  
 39 50.9 48.4 72.6





#57

1,3-Dichloropropane

Concen: 1.075 ug/l

RT: 9.482 min Scan# 1377

Delta R.T. 0.177 min

Lab File: VX045881.D

Acq: 22 Apr 2025 13:44

Instrument :

MSVOA\_X

ClientSampleId :





PB167661ZHE#02

Tgt Ion: 76 Resp: 1607

Ion Ratio Lower Upper

76 100

78 0.0 26.0 39.0#



#59

2-Hexanone

Concen: 13.176 ug/l

RT: 9.378 min Scan# 1360

Delta R.T. -0.049 min

Lab File: VX045881.D




Acq: 22 Apr 2025 13:44


Tgt Ion: 43 Resp: 14402

Ion Ratio Lower Upper

43 100

58 0.0 24.9 74.6#









#62  
4-Bromofluorobenzene  
Concen: 53.699 ug/l  
RT: 11.079 min Scan# 1639  
Delta R.T. 0.000 min  
Lab File: VX045881.D  
Acq: 22 Apr 2025 13:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
PB167661ZHE#02

Tgt Ion: 95 Resp: 58944  
Ion Ratio Lower Upper  
95 100  
174 69.3 0.0 135.8  
176 65.6 0.0 131.4




#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 10.049 min Scan# 1470  
Delta R.T. 0.000 min  
Lab File: VX045881.D  
Acq: 22 Apr 2025 13:44

Tgt Ion: 117 Resp: 115940  
Ion Ratio Lower Upper  
117 100  
82 63.7 49.2 73.8  
119 31.5 25.1 37.7











#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 12.018 min Scan# 11  
Delta R.T. 0.000 min  
Lab File: VX045881.D  
Acq: 22 Apr 2025 13:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
PB167661ZHE#02


Tgt Ion:152 Resp: 48065  
Ion Ratio Lower Upper  
152 100  
115 64.5 46.9 140.7  
150 155.0 0.0 349.4



#74  
N-ethyl acetate  
Concen: 0.218 ug/l  
RT: 10.896 min Scan# 1609  
Delta R.T. 0.055 min  
Lab File: VX045881.D  
Acq: 22 Apr 2025 13:44

Tgt Ion: 43 Resp: 401  
Ion Ratio Lower Upper  
43 100  
70 0.0 30.1 45.1#  
55 0.0 18.3 27.5#  
61 0.0 17.5 26.3#





#76

1,2,3-Trichloropropane

Concen: 29.097 ug/l

RT: 11.079 min Scan# 1639

Delta R.T. -0.158 min

Lab File: VX045881.D

Acq: 22 Apr 2025 13:44

Instrument :

MSVOA\_X

ClientSampleId :





PB167661ZHE#02

Tgt Ion: 75 Resp: 34114

Ion Ratio Lower Upper

75 100

77 1.1 20.0 59.9#



#81

trans-1,4-Dichloro-2-butene

Concen: 98.752 ug/l

RT: 11.079 min Scan# 1639

Delta R.T. 0.061 min

Lab File: VX045881.D

Acq: 22 Apr 2025 13:44




Tgt Ion: 75 Resp: 34114


Ion Ratio Lower Upper

75 100

53 0.0 101.8 152.6#

89 0.0 37.2 55.8#









#93  
1,2,4-Trichlorobenzene  
Concen: 0.319 ug/l  
RT: 13.597 min Scan# 2050  
Delta R.T. 0.012 min  
Lab File: VX045881.D  
Acq: 22 Apr 2025 13:44




Instrument :  
MSVOA\_X  
ClientSampleId :  
PB167661ZHE#02


Tgt Ion:180 Resp: 286  
Ion Ratio Lower Upper  
180 100  
182 119.9 47.9 143.6  
145 0.0 17.9 53.8#



#95  
Naphthalene  
Concen: 0.474 ug/l  
RT: 13.780 min Scan# 2082  
Delta R.T. 0.006 min  
Lab File: VX045881.D  
Acq: 22 Apr 2025 13:44

Tgt Ion:128 Resp: 1580  
Ion Ratio Lower Upper  
128 100  
127 9.2 10.3 15.5#  
129 11.1 8.6 12.8





#96

1,2,3-Trichlorobenzene

Concen: 0.304 ug/l

RT: 13.963 min Scan# 21

Delta R.T. 0.006 min

Lab File: VX045881.D


Acq: 22 Apr 2025 13:44

Instrument :

MSVOA\_X

ClientSampleId :

PB167661ZHE#02



Tgt Ion:180 Resp: 285

Ion Ratio Lower Upper

180 100

182 110.9 47.3 141.8

145 18.6 18.2 54.6

