

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042225\
 Data File : VX045882.D
 Acq On : 22 Apr 2025 14:07
 Operator : JC/MD
 Sample : PB167661ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB167661ZHE#03

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 04/23/2025
 Supervised By : Mahesh Dadoda 04/23/2025

Quant Time: Apr 23 01:34:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	64526	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	127180	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	120628	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	52334	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63029	53.414	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 106.820%			
35) Dibromofluoromethane	5.379	113	45986	50.963	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.920%			
50) Toluene-d8	8.647	98	161095	51.149	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 102.300%			
62) 4-Bromofluorobenzene	11.079	95	61836	53.901	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 107.800%			
Target Compounds					Qvalue	
16) Acetone	2.386	43	25891	53.433	ug/l	100
18) Methyl Acetate	2.709	43	2142	1.920	ug/l	95
20) Methylene Chloride	2.782	84	2963	3.266	ug/l	89
25) 2-Butanone	4.593	43	4030m	5.682	ug/l	
43) Isopropyl Acetate	6.348	43	11194	4.810	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042225\
Data File : VX045882.D
Acq On : 22 Apr 2025 14:07
Operator : JC/MD
Sample : PB167661ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB167661ZHE#03

Quant Time: Apr 23 01:34:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 04/23/2025
Supervised By :Mahesh Dadoda 04/23/2025

