

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042225\
 Data File : VX045884.D
 Acq On : 22 Apr 2025 14:54
 Operator : JC/MD
 Sample : PB167661ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB167661ZHE#05

Quant Time: Apr 23 01:34:36 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	67944	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	132966	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	126071	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	53730	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	65657	52.842	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.680%
35) Dibromofluoromethane	5.379	113	47409	50.254	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.500%
50) Toluene-d8	8.647	98	169325	51.423	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.840%
62) 4-Bromofluorobenzene	11.079	95	66170	55.169	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	110.340%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	26037	51.032	ug/l	99
18) Methyl Acetate	2.709	43	2172	1.849	ug/l #	90
20) Methylene Chloride	2.782	84	3472	3.635	ug/l	93
25) 2-Butanone	4.587	43	3804	5.093	ug/l	93
43) Isopropyl Acetate	6.348	43	10877	4.471	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042225\
Data File : VX045884.D
Acq On : 22 Apr 2025 14:54
Operator : JC/MD
Sample : PB167661ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB167661ZHE#05

Quant Time: Apr 23 01:34:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

