

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042225\
 Data File : VX045888.D
 Acq On : 22 Apr 2025 16:27
 Operator : JC/MD
 Sample : PB167661ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB167661ZHE#01

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 04/23/2025
 Supervised By : Mahesh Dadoda 04/23/2025

Quant Time: Apr 23 01:35:29 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	70819	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	136783	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	127546	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	52330	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	69371	53.564	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 107.120%			
35) Dibromofluoromethane	5.379	113	50274	51.803	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.600%			
50) Toluene-d8	8.647	98	172125	50.814	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 101.620%			
62) 4-Bromofluorobenzene	11.079	95	63195	51.218	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 102.440%			
Target Compounds						
16) Acetone	2.380	43	26393	49.629	ug/l	98
18) Methyl Acetate	2.703	43	2543	2.077	ug/l	95
20) Methylene Chloride	2.788	84	2655	2.667	ug/l	93
25) 2-Butanone	4.593	43	4563m	5.862	ug/l	
43) Isopropyl Acetate	6.342	43	12309	4.918	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042225\
Data File : VX045888.D
Acq On : 22 Apr 2025 16:27
Operator : JC/MD
Sample : PB167661ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB167661ZHE#01

Quant Time: Apr 23 01:35:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 04/23/2025
Supervised By :Mahesh Dadoda 04/23/2025

