

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042225\
 Data File : VX045890.D
 Acq On : 22 Apr 2025 17:13
 Operator : JC/MD
 Sample : Q1842-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PL-714-COMP-09

Quant Time: Apr 23 01:35:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	64241	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	129420	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	120391	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	51199	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64770	55.133	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.260%
35) Dibromofluoromethane	5.379	113	46466	50.604	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.200%
50) Toluene-d8	8.647	98	161474	50.382	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.760%
62) 4-Bromofluorobenzene	11.079	95	61134	52.367	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	104.740%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	22438	46.513	ug/l	99
18) Methyl Acetate	2.709	43	2540	2.287	ug/l	96
20) Methylene Chloride	2.782	84	3606	3.993	ug/l	92
43) Isopropyl Acetate	6.342	43	10763	4.545	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042225\
Data File : VX045890.D
Acq On : 22 Apr 2025 17:13
Operator : JC/MD
Sample : Q1842-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PL-714-COMP-09

Quant Time: Apr 23 01:35:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

