

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042225\
 Data File : VX045893.D
 Acq On : 22 Apr 2025 18:23
 Operator : JC/MD
 Sample : Q1842-30
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PL-714-COMP-12

Quant Time: Apr 23 01:36:29 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	65569	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	128493	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	121521	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	48784	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63993	53.368	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 106.740%		
35) Dibromofluoromethane	5.385	113	47215	51.790	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 103.580%		
50) Toluene-d8	8.647	98	163926	51.516	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 103.040%		
62) 4-Bromofluorobenzene	11.079	95	60788	52.446	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 104.900%		
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	23181	47.080	ug/l	99
18) Methyl Acetate	2.709	43	2670	2.355	ug/l #	88
20) Methylene Chloride	2.782	84	13160	14.277	ug/l	91
43) Isopropyl Acetate	6.348	43	10254	4.361	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042225\
Data File : VX045893.D
Acq On : 22 Apr 2025 18:23
Operator : JC/MD
Sample : Q1842-30
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PL-714-COMP-12

Quant Time: Apr 23 01:36:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

