

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042225\
 Data File : VX045895.D
 Acq On : 22 Apr 2025 19:10
 Operator : JC/MD
 Sample : Q1842-42
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PL-714-COMP-14

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 04/23/2025
 Supervised By : Mahesh Dadoda 04/23/2025

Quant Time: Apr 23 01:36:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	65143	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	129542	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	118970	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	49241	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64003	53.726	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 107.460%			
35) Dibromofluoromethane	5.385	113	46050	50.103	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.200%			
50) Toluene-d8	8.647	98	161665	50.394	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 100.780%			
62) 4-Bromofluorobenzene	11.079	95	59971	51.322	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 102.640%			
Target Compounds						
16) Acetone	2.379	43	33962	69.426	ug/l	97
18) Methyl Acetate	2.709	43	2620	2.326	ug/l	97
20) Methylene Chloride	2.782	84	208037	227.163	ug/l	97
25) 2-Butanone	4.586	43	4479m	6.255	ug/l	
43) Isopropyl Acetate	6.342	43	9217	3.889	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042225\
Data File : VX045895.D
Acq On : 22 Apr 2025 19:10
Operator : JC/MD
Sample : Q1842-42
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PL-714-COMP-14

Quant Time: Apr 23 01:36:52 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 04/23/2025
Supervised By :Mahesh Dadoda 04/23/2025

