



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042225\
Data File : VX045874.D
Acq On : 22 Apr 2025 08:27
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Title : SW846 8260
Last Update : Wed Apr 02 03:11:43 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.9	7294	PASS
75	95	30	60	56.7	18017	PASS
95	95	100	100	100.0	31800	PASS
96	95	5	9	6.6	2085	PASS
173	174	0.00	2	1.5	331	PASS
174	95	50	100	68.5	21787	PASS
175	174	5	9	6.7	1450	PASS
176	174	95	101	95.3	20760	PASS
177	176	5	9	6.2	1282	PASS