

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042319\
 Data File : VX009110.D
 Acq On : 23 Apr 2019 11:00
 Operator : JC/SP
 Sample : VX0423MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0423MBL01

Quant Time: Apr 24 05:08:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	254091	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	411820	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	353937	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	148459	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	168291	47.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.00%	
35) Dibromofluoromethane	5.50	113	124090	46.95	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.90%	
50) Toluene-d8	8.71	98	509463	48.33	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.66%	
62) 4-Bromofluorobenzene	11.13	95	169084	45.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.88%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042319\
Data File : VX009110.D
Acq On : 23 Apr 2019 11:00
Operator : JC/SP
Sample : VX0423MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0423MBL01

Quant Time: Apr 24 05:08:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

