

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042319\
 Data File : VX009115.D
 Acq On : 23 Apr 2019 12:56
 Operator : JC/SP
 Sample : K2537-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS042MW019-190422

Quant Time: Apr 24 05:17:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	244363	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	402209	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	344277	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	147043	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	164909	47.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.78%	
35) Dibromofluoromethane	5.50	113	121304	46.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.98%	
50) Toluene-d8	8.71	98	500775	48.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.30%	
62) 4-Bromofluorobenzene	11.13	95	164992	45.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.80%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	6865	2.163	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042319\
Data File : VX009115.D
Acq On : 23 Apr 2019 12:56
Operator : JC/SP
Sample : K2537-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS042MW019-190422

Quant Time: Apr 24 05:17:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

