

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042319\
 Data File : VX009117.D
 Acq On : 23 Apr 2019 13:43
 Operator : JC/SP
 Sample : K2537-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS039MW002-190422

Quant Time: Apr 24 05:24:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	239392	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	385777	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	337736	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	144267	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	156574	46.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.82%	
35) Dibromofluoromethane	5.50	113	116392	47.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.02%	
50) Toluene-d8	8.71	98	485273	49.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.30%	
62) 4-Bromofluorobenzene	11.13	95	159645	45.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.60%	
Target Compounds						
4) Vinyl Chloride	1.41	62	3430	1.041	ug/l	95
27) cis-1,2-Dichloroethene	4.60	96	10788	3.470	ug/l	93
44) Trichloroethene	7.21	130	12042	4.176	ug/l	93
64) Tetrachloroethene	9.34	164	41156	16.041	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042319\
Data File : VX009117.D
Acq On : 23 Apr 2019 13:43
Operator : JC/SP
Sample : K2537-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS039MW002-190422

Quant Time: Apr 24 05:24:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

