

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042319\
 Data File : VX009118.D
 Acq On : 23 Apr 2019 14:06
 Operator : JC/SP
 Sample : K2537-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY070MW233-190422

Quant Time: Apr 24 05:26:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	236508	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	387309	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	325330	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	132390	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	159494	47.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.70%	
35) Dibromofluoromethane	5.50	113	117045	47.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.18%	
50) Toluene-d8	8.71	98	474900	47.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.82%	
62) 4-Bromofluorobenzene	11.13	95	150789	43.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.18%	
Target Compounds						
4) Vinyl Chloride	1.41	62	62061	19.066	ug/l	89
21) trans-1,2-Dichloroethene	3.17	96	5166	1.910	ug/l	91
27) cis-1,2-Dichloroethene	4.60	96	7642	2.488	ug/l	90
52) Toluene	8.79	92	2432	0.343	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042319\
Data File : VX009118.D
Acq On : 23 Apr 2019 14:06
Operator : JC/SP
Sample : K2537-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY070MW233-190422

Quant Time: Apr 24 05:26:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

