

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042319\
 Data File : VX009122.D
 Acq On : 23 Apr 2019 15:40
 Operator : JC/SP
 Sample : K2535-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S37-0018

Quant Time: Apr 24 05:32:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	232830	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	382554	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	318110	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	129948	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	158557	48.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.64%	
35) Dibromofluoromethane	5.49	113	117073	47.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.36%	
50) Toluene-d8	8.71	98	468329	47.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.66%	
62) 4-Bromofluorobenzene	11.13	95	149400	43.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.46%	
Target Compounds						
16) Acetone	2.45	43	4119	2.421	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042319\
Data File : VX009122.D
Acq On : 23 Apr 2019 15:40
Operator : JC/SP
Sample : K2535-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S37-0018

Quant Time: Apr 24 05:32:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

