

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042319\
 Data File : VX009123.D
 Acq On : 23 Apr 2019 16:03
 Operator : JC/SP
 Sample : K2535-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S37-ER-2019-1

Quant Time: Apr 24 05:34:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	226975	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	367697	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	311336	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	129725	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	153301	47.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.86%	
35) Dibromofluoromethane	5.50	113	113545	48.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.24%	
50) Toluene-d8	8.71	98	454941	48.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.68%	
62) 4-Bromofluorobenzene	11.13	95	145024	43.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.30%	
Target Compounds						
16) Acetone	2.45	43	5752	3.468	ug/l	95
20) Methylene Chloride	2.86	84	38052	13.414	ug/l	98
29) Tetrahydrofuran	5.13	42	166999	94.102	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042319\
Data File : VX009123.D
Acq On : 23 Apr 2019 16:03
Operator : JC/SP
Sample : K2535-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S37-ER-2019-1

Quant Time: Apr 24 05:34:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

