

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042319\
 Data File : VX009124.D
 Acq On : 23 Apr 2019 16:27
 Operator : JC/SP
 Sample : K2537-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS042MW024-190422

Quant Time: Apr 24 05:36:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	231042	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	382023	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	327439	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	136332	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	156457	48.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.10%	
35) Dibromofluoromethane	5.50	113	115855	47.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.50%	
50) Toluene-d8	8.71	98	472858	48.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.72%	
62) 4-Bromofluorobenzene	11.13	95	153320	44.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.84%	
Target Compounds						
21) trans-1,2-Dichloroethene	3.17	96	3267	1.236	ug/l	99
27) cis-1,2-Dichloroethene	4.60	96	18679	6.225	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042319\
Data File : VX009124.D
Acq On : 23 Apr 2019 16:27
Operator : JC/SP
Sample : K2537-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS042MW024-190422

Quant Time: Apr 24 05:36:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

