

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042320\
 Data File : VX015896.D
 Acq On : 23 Apr 2020 17:24
 Operator : JC/SP
 Sample : L2403-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S73-0238

Quant Time: Apr 24 06:45:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042220W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 23 04:36:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	858340	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	1322759	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1318963	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	734715	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	483073	49.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.96%	
35) Dibromofluoromethane	5.47	113	392293	51.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.76%	
50) Toluene-d8	8.70	98	1638695	53.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.62%	
62) 4-Bromofluorobenzene	11.12	95	696701	55.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.50%	
Target Compounds						
4) Vinyl Chloride	1.39	62	51225	5.881	ug/l	99
27) cis-1,2-Dichloroethene	4.57	96	12516	1.322	ug/l	80
52) Toluene	8.76	92	7291	0.322	ug/l	91
67) Ethyl Benzene	10.23	91	9848	0.213	ug/l	99
73) Isopropylbenzene	11.00	105	16545	0.355	ug/l	99
78) n-propylbenzene	11.34	91	29868	0.559	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042320\
Data File : VX015896.D
Acq On : 23 Apr 2020 17:24
Operator : JC/SP
Sample : L2403-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S73-0238

Quant Time: Apr 24 06:45:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042220W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 23 04:36:12 2020
Response via : Initial Calibration

