

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042320\
 Data File : VX015899.D
 Acq On : 23 Apr 2020 18:32
 Operator : JC/SP
 Sample : L2403-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S73-ER-2020-1

Quant Time: Apr 24 07:28:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042220W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 23 04:36:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	853060	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1288087	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1288322	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	722048	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	470306	48.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.94%	
35) Dibromofluoromethane	5.47	113	366013	49.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.46%	
50) Toluene-d8	8.70	98	1585141	52.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.92%	
62) 4-Bromofluorobenzene	11.12	95	682405	55.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.16%	
Target Compounds						
16) Acetone	2.42	43	9857	2.333	ug/l #	87
17) Carbon Disulfide	2.55	76	13170	0.485	ug/l #	94
25) 2-Butanone	4.64	43	27036	4.049	ug/l	93
29) Tetrahydrofuran	5.09	42	167464	37.916	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042320\
Data File : VX015899.D
Acq On : 23 Apr 2020 18:32
Operator : JC/SP
Sample : L2403-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S73-ER-2020-1

Quant Time: Apr 24 07:28:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042220W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 23 04:36:12 2020
Response via : Initial Calibration

