

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042320\
 Data File : VX015914.D
 Acq On : 24 Apr 2020 00:57
 Operator : JC/SP
 Sample : L2364-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 30963

Quant Time: Apr 24 08:19:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042220W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 23 04:36:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	845037	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1282800	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1267804	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	724001	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	475749	49.50	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.00%	
35) Dibromofluoromethane	5.47	113	373680	50.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.94%	
50) Toluene-d8	8.70	98	1551626	52.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.10%	
62) 4-Bromofluorobenzene	11.12	95	660414	54.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.02%	
Target Compounds						
16) Acetone	2.42	43	67283	16.078	ug/l	95
20) Methylene Chloride	2.84	84	48188	5.672	ug/l	96
43) Isopropyl Acetate	6.42	43	268869	11.777	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042320\
Data File : VX015914.D
Acq On : 24 Apr 2020 00:57
Operator : JC/SP
Sample : L2364-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
30963

Quant Time: Apr 24 08:19:57 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042220W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 23 04:36:12 2020
Response via : Initial Calibration

