

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042320\
 Data File : VX015917.D
 Acq On : 24 Apr 2020 02:05
 Operator : JC/SP
 Sample : L2378-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-6-S

Quant Time: Apr 24 08:26:50 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042220W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 23 04:36:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	858224	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1297090	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1283779	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	737713	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	474583	48.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.24%	
35) Dibromofluoromethane	5.47	113	379013	50.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.24%	
50) Toluene-d8	8.70	98	1558530	51.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.42%	
62) 4-Bromofluorobenzene	11.12	95	662000	53.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.08%	
Target Compounds						
16) Acetone	2.42	43	56933	13.395	ug/l	95
20) Methylene Chloride	2.84	84	42055	4.874	ug/l	98
43) Isopropyl Acetate	6.41	43	235767	10.213	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042320\
Data File : VX015917.D
Acq On : 24 Apr 2020 02:05
Operator : JC/SP
Sample : L2378-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-6-S

Quant Time: Apr 24 08:26:50 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042220W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 23 04:36:12 2020
Response via : Initial Calibration

