


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX042320\
 Data File : VX015903.D
 Acq On : 23 Apr 2020 20:25
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042220W.M
 Title : SW846 8260
 Last Update : Thu Apr 23 04:36:12 2020



AutoFind: Scans 1678, 1679, 1680; Background Corrected with Scan 1670

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.1	40466	PASS
75	95	30	60	48.3	108275	PASS
95	95	100	100	100.0	224170	PASS
96	95	5	9	6.4	14458	PASS
173	174	0.00	2	0.6	1185	PASS
174	95	50	100	85.0	190540	PASS
175	174	5	9	7.5	14316	PASS
176	174	95	101	98.0	186774	PASS
177	176	5	9	6.2	11595	PASS