

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042321\
 Data File : VX021840.D
 Acq On : 23 Apr 2021 14:28
 Operator : JC/MD
 Sample : M2138-03DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-DDC-2-PDDL

Quant Time: Apr 26 04:03:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 20 13:46:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	177466	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	291773	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	258951	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	119634	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	95996	47.68	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.36%
35) Dibromofluoromethane	5.464	113	92520	49.58	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.16%
50) Toluene-d8	8.695	98	344556	48.03	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.06%
62) 4-Bromofluorobenzene	11.116	95	120838	45.10	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.20%
Target Compounds						
64) Tetrachloroethene	9.317	164	194402	96.01	ug/l	98
90) Hexachloroethane	12.579	117	4335	3.08	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042321\
Data File : VX021840.D
Acq On : 23 Apr 2021 14:28
Operator : JC/MD
Sample : M2138-03DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-DDC-2-PDDL

Quant Time: Apr 26 04:03:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 20 13:46:22 2021
Response via : Initial Calibration

