

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042321\
 Data File : VX021841.D
 Acq On : 23 Apr 2021 14:51
 Operator : JC/MD
 Sample : M2134-01DL 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 210421084-02-VOADL

Quant Time: Apr 26 04:03:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 20 13:46:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	173848	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	293588	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	263935	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	127526	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.032	65	97575	49.48	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	98.96%		
35) Dibromofluoromethane	5.465	113	93241	49.66	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.32%		
50) Toluene-d8	8.696	98	347238	48.10	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	96.20%		
62) 4-Bromofluorobenzene	11.122	95	124486	46.18	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	92.36%		
Target Compounds						Qvalue
11) Tert butyl alcohol	3.008	59	28211	65.32	ug/l	96
16) Acetone	2.422	43	10975	16.19	ug/l	95
25) 2-Butanone	4.635	43	21018	18.31	ug/l	98
29) Tetrahydrofuran	5.093	42	5989	8.02	ug/l	93
95) Naphthalene	13.817	128	6159	0.72	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042321\
Data File : VX021841.D
Acq On : 23 Apr 2021 14:51
Operator : JC/MD
Sample : M2134-01DL 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
210421084-02-VOADL

Quant Time: Apr 26 04:03:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 20 13:46:22 2021
Response via : Initial Calibration

