

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042321\
 Data File : VX021849.D
 Acq On : 23 Apr 2021 17:53
 Operator : JC/MD
 Sample : M2164-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW8-GW-278-280

Quant Time: Apr 26 04:06:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 20 13:46:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	173790	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	290759	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.098	117	258705	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	121623	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	94633	48.00	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	96.00%		
35) Dibromofluoromethane	5.464	113	92938	49.98	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.96%		
50) Toluene-d8	8.696	98	343920	48.10	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	96.20%		
62) 4-Bromofluorobenzene	11.116	95	122252	45.79	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	91.58%		
Target Compounds					Qvalue	
16) Acetone	2.422	43	8226	12.14	ug/l	93
25) 2-Butanone	4.641	43	2632	2.29	ug/l	91
52) Toluene	8.769	92	1308	0.26	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042321\
Data File : VX021849.D
Acq On : 23 Apr 2021 17:53
Operator : JC/MD
Sample : M2164-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-278-280

Quant Time: Apr 26 04:06:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 20 13:46:22 2021
Response via : Initial Calibration

