

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042321\
 Data File : VX021855.D
 Acq On : 23 Apr 2021 20:10
 Operator : JC/MD
 Sample : M2150-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TWP-20

Quant Time: Apr 26 04:07:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 20 13:46:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	177418	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	298475	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.097	117	272854	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	140838	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	98571	48.97	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	97.94%		
35) Dibromofluoromethane	5.470	113	93903	49.20	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	98.40%		
50) Toluene-d8	8.695	98	354249	48.27	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	96.54%		
62) 4-Bromofluorobenzene	11.122	95	130295	47.54	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	95.08%		
Target Compounds					Qvalue	
16) Acetone	2.422	43	9214	13.32	ug/l	98
20) Methylene Chloride	2.830	84	1106	0.53	ug/l #	72
25) 2-Butanone	4.641	43	3617	3.09	ug/l	91
65) Chlorobenzene	10.122	112	4402	0.78	ug/l	96
95) Naphthalene	13.816	128	78698	8.29	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042321\
Data File : VX021855.D
Acq On : 23 Apr 2021 20:10
Operator : JC/MD
Sample : M2150-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TWP-20

Quant Time: Apr 26 04:07:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 20 13:46:22 2021
Response via : Initial Calibration

