

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042321\
 Data File : VX021856.D
 Acq On : 23 Apr 2021 20:32
 Operator : JC/MD
 Sample : M2150-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TWP-40

Quant Time: Apr 26 04:07:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 20 13:46:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	177481	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	299279	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.098	117	271386	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	135426	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	98278	48.81	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.62%
35) Dibromofluoromethane	5.464	113	95139	49.71	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.42%
50) Toluene-d8	8.695	98	353308	48.01	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.02%
62) 4-Bromofluorobenzene	11.122	95	128018	46.58	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.16%
Target Compounds						
16) Acetone	2.416	43	2556	3.69	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042321\
Data File : VX021856.D
Acq On : 23 Apr 2021 20:32
Operator : JC/MD
Sample : M2150-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TWP-40

Quant Time: Apr 26 04:07:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 20 13:46:22 2021
Response via : Initial Calibration

