

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042321\
 Data File : VX021860.D
 Acq On : 23 Apr 2021 22:03
 Operator : JC/MD
 Sample : M2138-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-MW-15S

Quant Time: Apr 26 04:08:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 20 13:46:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	180696	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	304033	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	275165	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	139793	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	99629	48.60	ug/l	0.00
Spiked Amount 50.000	Range	78 - 117	Recovery	=	97.20%	
35) Dibromofluoromethane	5.464	113	96273	49.52	ug/l	0.00
Spiked Amount 50.000	Range	75 - 124	Recovery	=	99.04%	
50) Toluene-d8	8.695	98	359339	48.07	ug/l	0.00
Spiked Amount 50.000	Range	92 - 112	Recovery	=	96.14%	
62) 4-Bromofluorobenzene	11.122	95	132027	47.29	ug/l	0.00
Spiked Amount 50.000	Range	83 - 123	Recovery	=	94.58%	
Target Compounds						Qvalue
16) Acetone	2.422	43	1964	2.79	ug/l	88
30) Chloroform	5.178	83	2268	0.66	ug/l	89
64) Tetrachloroethene	9.317	164	68179	31.69	ug/l	97
90) Hexachloroethane	12.579	117	1562	0.95	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042321\
Data File : VX021860.D
Acq On : 23 Apr 2021 22:03
Operator : JC/MD
Sample : M2138-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-MW-15S

Quant Time: Apr 26 04:08:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 20 13:46:22 2021
Response via : Initial Calibration

