

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042324\
 Data File : VX041214.D
 Acq On : 23 Apr 2024 14:13
 Operator : JC/MD
 Sample : P2233-01DL 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RMW-03B-31-042224DL

Quant Time: Apr 24 01:47:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042224W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 23 05:30:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	193100	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	246002	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	228916	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	120427	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	99054	53.888	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.780%	
35) Dibromofluoromethane	5.385	113	87510	58.480	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 116.960%	
50) Toluene-d8	8.647	98	322362	63.851	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 127.700%#	
62) 4-Bromofluorobenzene	11.079	95	111139	59.641	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 119.280%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.306	96	735	0.522	ug/l	85
16) Acetone	2.392	43	1336	1.997	ug/l	# 79
20) Methylene Chloride	2.788	84	1785	1.092	ug/l	86
24) 1,1-Dichloroethane	3.605	63	6808	2.464	ug/l	98
27) cis-1,2-Dichloroethene	4.495	96	1570	0.814	ug/l	80
30) Chloroform	5.099	83	4753	1.603	ug/l	95
44) Trichloroethene	7.123	130	59478	34.597	ug/l	84
52) Toluene	8.720	92	9881	2.652	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042324\
Data File : VX041214.D
Acq On : 23 Apr 2024 14:13
Operator : JC/MD
Sample : P2233-01DL 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RMW-03B-31-042224DL

Quant Time: Apr 24 01:47:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042224W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 23 05:30:28 2024
Response via : Initial Calibration

