

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042324\
 Data File : VX041223.D
 Acq On : 23 Apr 2024 18:22
 Operator : JC/MD
 Sample : P2201-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Apr 24 01:51:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042224W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 23 05:30:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	207913	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	303592	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	277981	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	116660	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	106825	53.975	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.940%	
35) Dibromofluoromethane	5.385	113	96002	51.985	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.980%	
50) Toluene-d8	8.647	98	304004	48.792	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 97.580%	
62) 4-Bromofluorobenzene	11.079	95	100817	43.839	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 87.680%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042324\
Data File : VX041223.D
Acq On : 23 Apr 2024 18:22
Operator : JC/MD
Sample : P2201-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Apr 24 01:51:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042224W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 23 05:30:28 2024
Response via : Initial Calibration

