

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042325\
 Data File : VX045911.D
 Acq On : 23 Apr 2025 16:29
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Apr 24 02:15:48 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	66779	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	131820	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	120312	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	48447	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64525	52.837	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 105.680%	
35) Dibromofluoromethane	5.385	113	47462	50.747	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.500%	
50) Toluene-d8	8.647	98	163657	50.133	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.260%	
62) 4-Bromofluorobenzene	11.079	95	60449	50.837	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 101.680%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042325\
Data File : VX045911.D
Acq On : 23 Apr 2025 16:29
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Apr 24 02:15:48 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 731
Delta R.T. -0.006 min
Lab File: VX045911.D
Acq: 23 Apr 2025 16:29




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:168 Resp: 66779
Ion Ratio Lower Upper
168 100
99 68.3 52.3 78.5



#33
1,2-Dichloroethane-d4
Concen: 52.837 ug/l
RT: 5.952 min Scan# 798
Delta R.T. -0.000 min
Lab File: VX045911.D
Acq: 23 Apr 2025 16:29

Tgt Ion: 65 Resp: 64525
Ion Ratio Lower Upper
65 100
67 48.3 0.0 99.0









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 911
Delta R.T. -0.000 min
Lab File: VX045911.D
Acq: 23 Apr 2025 16:29




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:114 Resp: 131820
Ion Ratio Lower Upper
114 100
63 23.6 0.0 46.8
88 17.2 0.0 35.4



#35
Dibromofluoromethane
Concen: 50.747 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.006 min
Lab File: VX045911.D
Acq: 23 Apr 2025 16:29

Tgt Ion:113 Resp: 47462
Ion Ratio Lower Upper
113 100
111 101.6 81.8 122.6
192 16.6 13.8 20.6









#50
Toluene-d8
Concen: 50.133 ug/l
RT: 8.647 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX045911.D
Acq: 23 Apr 2025 16:29




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion: 98 Resp: 163657
Ion Ratio Lower Upper
98 100
100 64.2 52.2 78.4



#62
4-Bromofluorobenzene
Concen: 50.837 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX045911.D
Acq: 23 Apr 2025 16:29

Tgt Ion: 95 Resp: 60449
Ion Ratio Lower Upper
95 100
174 66.8 0.0 135.8
176 64.0 0.0 131.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX045911.D
Acq: 23 Apr 2025 16:29

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Tgt Ion:117 Resp: 120312
Ion Ratio Lower Upper
117 100
82 62.1 49.2 73.8
119 31.5 25.1 37.7



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX045911.D
Acq: 23 Apr 2025 16:29

Tgt Ion:152 Resp: 48447
Ion Ratio Lower Upper
152 100
115 64.6 46.9 140.7
150 159.4 0.0 349.4

