

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042419\
 Data File : VX009140.D
 Acq On : 24 Apr 2019 13:14
 Operator : JC/SP
 Sample : K2535-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S37-0018

Quant Time: Apr 25 06:44:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	229319	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	371200	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	311926	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	126173	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	153617	47.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.06%	
35) Dibromofluoromethane	5.50	113	113049	47.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.90%	
50) Toluene-d8	8.71	98	456229	48.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.04%	
62) 4-Bromofluorobenzene	11.13	95	143963	42.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.86%	
Target Compounds						
16) Acetone	2.45	43	6890	4.111	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042419\
Data File : VX009140.D
Acq On : 24 Apr 2019 13:14
Operator : JC/SP
Sample : K2535-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S37-0018

Quant Time: Apr 25 06:44:49 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

