

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042419\
 Data File : VX009141.D
 Acq On : 24 Apr 2019 13:37
 Operator : JC/SP
 Sample : K2535-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S37-ER-2019-1

Quant Time: Apr 25 06:47:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	229612	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	373007	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	314516	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	131058	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	154472	47.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.48%	
35) Dibromofluoromethane	5.50	113	112469	46.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.96%	
50) Toluene-d8	8.71	98	458599	48.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.08%	
62) 4-Bromofluorobenzene	11.13	95	149270	44.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.58%	
Target Compounds						
16) Acetone	2.45	43	6505	3.877	ug/l	93
20) Methylene Chloride	2.86	84	38009	13.245	ug/l	94
29) Tetrahydrofuran	5.13	42	167207	93.137	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042419\
Data File : VX009141.D
Acq On : 24 Apr 2019 13:37
Operator : JC/SP
Sample : K2535-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S37-ER-2019-1

Quant Time: Apr 25 06:47:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

