

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042419\
 Data File : VX009147.D
 Acq On : 24 Apr 2019 17:38
 Operator : JC/SP
 Sample : K2555-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Apr 25 07:14:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	204941	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	6.85	114	330714	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	278982	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	112994	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	136444	47.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.48%	
35) Dibromofluoromethane	5.49	113	101081	47.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.24%	
50) Toluene-d8	8.71	98	406381	48.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.02%	
62) 4-Bromofluorobenzene	11.13	95	129046	43.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.38%	
Target Compounds						
16) Acetone	2.44	43	4867	3.250	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042419\
Data File : VX009147.D
Acq On : 24 Apr 2019 17:38
Operator : JC/SP
Sample : K2555-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Apr 25 07:14:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

