

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042419\
 Data File : VX009149.D
 Acq On : 24 Apr 2019 18:25
 Operator : JC/SP
 Sample : K2555-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 HW-23

Quant Time: Apr 25 07:20:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	220444	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	360256	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	303664	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	125637	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	149416	48.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.20%	
35) Dibromofluoromethane	5.50	113	110761	47.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.82%	
50) Toluene-d8	8.71	98	443128	48.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.12%	
62) 4-Bromofluorobenzene	11.13	95	141459	43.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.92%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042419\
Data File : VX009149.D
Acq On : 24 Apr 2019 18:25
Operator : JC/SP
Sample : K2555-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
HW-23

Quant Time: Apr 25 07:20:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

